

October 11, 2012

CERTIFICATE OF CONFORMANCE

We confirm that the device supplied (7314X Series) complies with has the following approvals: IEC 6060101-1-2; CSA-C22.2#601.1; UL60601-1-2; ISO 10079-1; IEC 60601-1; IEC 60529 IP12: IEC 60601-1-6 and EN 60601-1. The 7314X Series is manufactured under the good manufacturing practices regulations specified by the United States Food and Drug Administration and quality standard ISO 13485:2003 (Registration TUV – 44 232 08 358109-001). The device meets the functional product specification of DeVilbiss Healthcare.

George Illar Quality Manager Devilbiss Healthcare 100 DeVilbiss Dr. Somerset, PA 15501

