

## EU-CONFORMITY DECLARATION

It is hereby declared that the "**Laerdal Suction Unit Wall Bracket, models 782600 - 782699**" with accessories, manufactured by Laerdal Medical AS, P.O.Box 377, N-4002 Stavanger, is manufactured in accordance with the product described in the technical file.

Conformity is declared for the essential requirements of Council Directive **93/42/EEC** regarding the harmonization of the membership's laws on **Medical Devices**.


The declaration of conformity is based on the following technical standards:

**IEC60601**  
**EN1789**

-and all referenced standards at a lower level

The product is in risk-class I according to rule **1** in appendix IX of the Directive.

For Laerdal Medical AS

  
-----

Pål Frafjord  
Technical Product Manager

*01.06.2005*  
-----

Date



Laerdal Medical AS  
P.O.Box 377  
N-4002 Stavanger  
Telephone: 51511700 Fax: 51511788